



MAXIMUM STORAGE CAPACITY
 1.2 & 2.0 ml VALS IN RACKS _____ 10,400
 NUMBER OF 13-2 RACKS (100 VAL) _____ 7
 NUMBER OF 13-2 RACKS (25 VAL) _____ 4
 TOTAL NUMBER OF 13-2 RACKS _____ 11
 NUMBER OF BLOOD BAGS STORED _____ 192
 SUC-1 CANISTER (2.5" X 2.5" X 11") _____ 36
 1.2 AND 2.0 ml PLASTIC VALS ON CANES _____ 6,048
 1/2 cc STRAWS (10/CANE) _____ 11,160

PERFORMANCE
 LN2 CAPACITY _____ 159 LITERS
 STATIC EVAPORATION RATE _____ 6.0 LITERS/DAY
 POWER SUPPLY (VARIABLE W/110 V/200 V TRANSFORMER) -24 VOLTS AC

UNIT DIMENSIONS
 NECK OPENING _____ 20.75 in./527mm
 USABLE HEIGHT _____ 30.9 in./784 mm
 OVERALL HEIGHT _____ 52.29 in / 1328 mm MAX
 OUTSIDE DIMENSIONS _____ 35.60 in x 31.22 in/904 mm x 793 mm
 INTERNAL DIAMETER _____ 20.75 in./527mm
 WEIGHT EMPTY _____ 213lbs./96.5 kg
 WEIGHT FULL _____ 496lbs./225.5kg

REV	APPV	ECR	REVISION DESCRIPTION	BY	DATE
B	JB		REVISED SPECIFICATIONS	JSB	2/6/97
A		201,253	REV DIM AND ADDED FRONT VIEW	LF	8-27-96

THE MATERIALS AND INFORMATION, INCLUDING THE PRINCIPLES OF DESIGN PRESENTED BY THIS PRINT, IS THE EXCLUSIVE PROPERTY OF MINNESOTA VALLEY ENGINEERING, AND IS CONFIDENTIAL INFORMATION. ACCORDINGLY, THIS INFORMATION IS SUBMITTED TO YOU WITH THE AGREEMENT THAT IT IS NOT TO BE REPRODUCED, COPIED, OR LOANED, IN PART OR IN WHOLE, NOR IS THE INFORMATION TO BE RELAYED TO ANY COMPANY. ACCEPTANCE OF THIS DRAWING WILL BE CONSTRUED AS AN AGREEMENT TO THE ABOVE.

DESIGNATES A NOTE REFERENCE.

MAT'L

FINISH P/N
10779311

APP'VD	DATE	USED ON
DRAWN BY JBP	11-20-95	201,337
CHECKED BY JB	2-6-97	ORIGINAL ECR/ER

TITLE:
 MVE INC.
 CANTON, GEORGIA 30114

XLC-511, SPEC

DRAWING NO.	REV.
B-10779311	B

SCALE: NONE DO NOT SCALE DRAWING SHEET: 1 of 1